

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0509N

System: GNDRAD PIR

Site: Nauru

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PIR	31311	Gndrad	9/26/2004	-10.57 K0 .26064 K1 1.013 K2 -3.227 K3	na	9/19/2005	20:30	22:40

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PIR	31296	Gndrad	9/19/2005	-16.51 K0 .25324 K1 1.032 K2 (-2.648) K3	na	X

Replacement completed by: Troy Culgan

## NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: